

**Antibiotic Susceptibility of *Enterococcus faecium* ATCC® BAA-2320™**

<b>Antimicrobial</b>	<b>MIC<sup>a</sup></b>	<b>Interpretation<sup>b</sup></b>	<b>Antimicrobial</b>	<b>MIC</b>	<b>Interpretation</b>
Benzympenicillin	≥64	R	Erythromycin	≥8	R
Ampicillin	≥32	R	Quinupristin/Dalfopristin	0.5	S
Ciprofloxacin	≥8	R	Linezolid	2	S
Levofloxacin	≥8	R	Vancomycin	≥32	R
			Tetracycline	≥16	R
			Tigecycline	≤0.12	S
			Nitrofurantoin	128	R

<sup>a</sup> Antibiotic susceptibility was obtained using a Vitek2 AST-GP67

<sup>b</sup> Parameter set: MIC Interpretation Guideline: CLSI M100ED33 (2023)  
Therapeutic Interpretation Guideline: Natural Resistance