



# Quantitative Genomic DNA from *Streptococcus pyogenes* strain SF370; M1 GAS

700294DQ™

## Description

ATCC® Genuine Nucleics can be used for assay development, verification, validation, monitoring of day-to-day test variation and lot-to-lot performance of molecular-based assays. The quantitative format allows for the generation of a standard curve for quantitative PCR (qPCR) to determine bacterial load and for other molecular biology applications.

**Organism:** *Streptococcus pyogenes* Rosenbach

**Derived from:** *Streptococcus pyogenes* SF370; M1 GAS (ATCC 700294)

**Genome sequenced strain:** Yes

**Type strain:** No

**Genotype:** lysogenic for a bacteriophage carrying pyrogenic exotoxin C (speC+) but negative for erythrogenic toxin A (speA-)

**Specification range:**  $\geq 1 \times 10^5$  copies/ $\mu$ L

**Volume:** 100  $\mu$ L

---

## Storage Conditions

**Product format:** Frozen

**Storage conditions:** -20°C or colder

---

## Intended Use

This product is intended for laboratory research use only. It is not intended for any animal or human therapeutic use, any human or animal consumption, or any diagnostic use.

# Quantitative Genomic DNA from *Streptococcus pyogenes* strain SF370; M1 GAS

700294DQ

BSL 1

ATCC determines the biosafety level of a material based on our risk assessment as guided by the current edition of *Biosafety in Microbiological and Biomedical Laboratories (BMBL)*, U.S. Department of Health and Human Services. It is your responsibility to understand the hazards associated with the material per your organization's policies and procedures as well as any other applicable regulations as enforced by your local or national agencies.

---

## Certificate of Analysis

For batch-specific test results, refer to the applicable certificate of analysis that can be found at [www.atcc.org](http://www.atcc.org).

---

## Handling Procedures

1. Thaw the vial at room temperature and immediately place on ice. Avoid exposing the DNA to repeated freeze-thaw cycles as it may result in degradation.
  2. Gently mix the sample to ensure an even distribution of material.
  3. Briefly centrifuge the tube before opening to ensure all the liquid is at the bottom.
- 

## Notes

Aliquoting is highly recommended to avoid multiple freeze-thaws.

---

## Material Citation

If use of this material results in a scientific publication, please cite the material in the following manner: Quantitative Genomic DNA from *Streptococcus pyogenes* strain

# Quantitative Genomic DNA from *Streptococcus pyogenes* strain SF370; M1 GAS

700294DQ

SF370; M1 GAS (ATCC 700294DQ)

---

## References

References and other information relating to this material are available at [www.atcc.org](http://www.atcc.org).

---

## Warranty

The product is provided 'AS IS' and the viability of ATCC® products is warranted for 30 days from the date of shipment, provided that the customer has stored and handled the product according to the information included on the product information sheet, website, and Certificate of Analysis. For living cultures, ATCC lists the media formulation and reagents that have been found to be effective for the product. While other unspecified media and reagents may also produce satisfactory results, a change in the ATCC and/or depositor-recommended protocols may affect the recovery, growth, and/or function of the product. If an alternative medium formulation or reagent is used, the ATCC warranty for viability is no longer valid. Except as expressly set forth herein, no other warranties of any kind are provided, express or implied, including, but not limited to, any implied warranties of merchantability, fitness for a particular purpose, manufacture according to cGMP standards, typicality, safety, accuracy, and/or noninfringement.

---

## Disclaimers

This product is intended for laboratory research use only. It is not intended for any animal or human therapeutic use, any human or animal consumption, or any diagnostic use. Any proposed commercial use is prohibited without a license from ATCC.

While ATCC uses reasonable efforts to include accurate and up-to-date information

## Quantitative Genomic DNA from *Streptococcus pyogenes* strain SF370; M1 GAS

700294DQ

on this product sheet, ATCC makes no warranties or representations as to its accuracy. Citations from scientific literature and patents are provided for informational purposes only. ATCC does not warrant that such information has been confirmed to be accurate or complete and the customer bears the sole responsibility of confirming the accuracy and completeness of any such information.

This product is sent on the condition that the customer is responsible for and assumes all risk and responsibility in connection with the receipt, handling, storage, disposal, and use of the ATCC product including without limitation taking all appropriate safety and handling precautions to minimize health or environmental risk. As a condition of receiving the material, the customer agrees that any activity undertaken with the ATCC product and any progeny or modifications will be conducted in compliance with all applicable laws, regulations, and guidelines. This product is provided 'AS IS' with no representations or warranties whatsoever except as expressly set forth herein and in no event shall ATCC, its parents, subsidiaries, directors, officers, agents, employees, assigns, successors, and affiliates be liable for indirect, special, incidental, or consequential damages of any kind in connection with or arising out of the customer's use of the product. While reasonable effort is made to ensure authenticity and reliability of materials on deposit, ATCC is not liable for damages arising from the misidentification or misrepresentation of such materials.

Please see the material transfer agreement (MTA) for further details regarding the use of this product. The MTA is available at [www.atcc.org](http://www.atcc.org).

---

### Copyright and Trademark Information

© ATCC 2023. All rights reserved.

ATCC is a registered trademark of the American Type Culture Collection.

---

### Revision

This information on this document was last updated on 2022-10-09

---

### Contact Information

## Quantitative Genomic DNA from *Streptococcus pyogenes* strain SF370; M1 GAS

700294DQ

ATCC

10801 University Boulevard

Manassas, VA 20110-2209

USA

US telephone: 800-638-6597

Worldwide telephone: +1-703-365-2700

Email: [tech@atcc.org](mailto:tech@atcc.org) or contact your local distributor

---